







2 off Pos.2 for cables
with Thickness > 4mm



POS. NO.	DESCRIPTION	MATERIAL	QTY
1	housing top part SUB-D 15 pin	Zinc alloy / Nickel plated	1
2	screw M3x8	Stainless Steel	4
3	Viton tube D=3, L=15mm	VITON	1
4	screw M3 x 6	Stainless Steel	2
5	cable clamp 15	Steel Nickel plated	2
6	screw 4-40 UNC	1.4305 (303)	2
7	housing bottom part SUB-D 15 pin	Zinc alloy / Nickel plated	1
8	viton tube 2x3.5	VITON	2



no leakcheck required Definition Leckprüfung (definiton leak check)				Kanten (EDGES)		PROJECTION	
Bemerkung (REMARKS)							
Toleranz (TOLERANCE)							
Maßstab (SCALE): 1:1		Format. (FORM) A4	PAGE NO 1	PAGES 1	Rauheit (ROUGHNESS) 		
Benennung (DESCRIPTION): housing for HV-SUB-D connector 15 pin strain relief							
Zeich. Nr. (Drw. No.): 211-HSG-D15-SR-V3						REV 02	
Art. Nr. (PART NO.) 211-HSG-D15-SR-V3							

Zeichnung Nutzung (SCOPE OF DRAWING)		Prüfung Zubehör. (CON + HSG CHECK)	
Drawing external usage for customer			
All informations provided in this drawing is considered to be confidential, proprietary and secret information. It is intended only for the use of individual or entity to whom it is sent. Individual or entity agrees to not disseminate or copy this information to anyone other than originally specified without the written consent Allectra GmbH.		Werkstoff (MATERIAL): Gewicht (WEIGHT):	
		g	
Oberfläche (SURFACE)		Maßstab (SCALE):	
Cleaned UHV- Compatible		1:1	
origin	name	DATE	NAME
	(drw/doc/sketch)	19.11.2015	FW
	date	24.07.2025	AD
		25.07.2025	JB
		25.07.2025	AD
02 Pos.2 Qty 2 to 4 changed		alllectra	
REV Beschreibung (CHANGING)		gesp. am (SAVE DAT.): 25.07.2025	