

01

connecting view
connector

Con1	Cable
1	1
19	2
10	3
2	4
20	5
11	6
3	7
21	8
12	9
4	10
22	11
13	12
5	13
23	14
14	15
6	16
24	17
15	18
7	19
25	20
16	21
8	22
26	23
17	24
9	25
18	26




01

connecting view
cable

Cable	Con1
1	1
2	19
3	10
4	2
5	20
6	11
7	3
8	21
9	12
10	4
11	22
12	13
13	5
14	23
15	14
16	6
17	24
18	15
19	7
20	25
21	16
22	8
23	26
24	17
25	9
26	18

cable

Lx +0/-20





connector 1 /211-FS26H-PK basic for feedthrough

POS-NR.	DESCRIPTION	QTY
1	UHV connector 26 Pin female	1
2	ribbon cable 26 pin Kapton insulated	1

no leakcheck required

Definition Leckprüfung (definition leak check)

Zeichnung Nutzung (SCOPE OF DRAWING) Drawing external usage for customer			Prüfung Zubehör. (CON + HSG CHECK)			Bemerkung (REMARKS)			Kanten (EDGES) 		
All informations provided in this drawing is considered to be confidential, proprietary and secret information. It is intended only for the use of individual or entity to whom it is sent. Individually or entity agrees to not disseminate or copy this information to anyone other than originally specified without the written consent Allectra GmbH.			Werkstoff (MATERIAL):		Gewicht (WEIGHT): g		Toleranz (TOLERANCE)				
			Oberfläche (SURFACE)		Maßstab (SCALE): 1:1		Format. (FORM) A4	PAGE NO 1	PAGES 1	Rauheit (ROUGHNESS) 	
origin	name Allectra standard			DATE	NAME		Benennung (DESCRIPTION): cable with 1 of 26 pin female connectors PEEK 1x 26 way ribbon cable				
	(drw/doc/sketch)		Design	20.06.2019	AD						
			Drawn	11.07.2025	JB						
	date		Check	14.07.2025	AD						
	01	connection sorted		Approv	14.07.2025	JB	Zeich. Nr. (Drw. No.): 380-HD26-FXPR-X				
											
REV	Beschreibung (CHANGING)		gesp. am (SAVE DAT.): 14.07.2025			Art. Nr. (PART NO.)					